- (iii) Within 5 days if required by \$803.53.
- (c) Information that reasonably suggests a reportable event occurred. (1) Information that reasonably suggests that a device has or may have caused or contributed to an MDR reportable event (i.e., death, serious injury, and, for manufacturers, a malfunction that would be likely to cause or contribute to a death or serious injury if the malfunction were to recur) includes any information, such as professional, scientific or medical facts and observations or opinions, that would reasonably suggest that a device has caused or may have caused or contributed to a reportable event.

(2) Entities required to report under this part do not have to report adverse events for which there is information that would cause a person who is qualified to make a medical judgment (e.g., a physician, nurse, risk manager, or biomedical engineer) to reach a reasonable conclusion that a device did not cause or contribute to a death or serious injury, or that a malfunction would not be likely to cause or contribute to a death or serious injury if it were to recur. Information which leads the qualified person to determine that a device-related event is or is not reportable must be contained in the MDR event files, as described in §803.18.

§803.21 Reporting codes.

(a) FDA has developed a MEDWATCH Mandatory Reporting Form Coding Manual for use with medical device reports. This manual contains codes for hundreds of adverse events for use with FDA Form 3500A. The coding manual is available from the Division of Small Manufacturer Assistance, Center for Devices and Radiological Health, 1350 Piccard Dr., Rockville, MD 20850, FAX 301-443-8818.

(b) FDA may use additional coding of information on the reporting forms or modify the existing codes on an ad hoc or generic basis. In such cases, FDA will ensure that the new coding information is available to all reporters.

§803.22 When not to file.

(a) Only one medical device report from the user facility or manufacturer is required under this part if the reporting entity becomes aware of information from multiple sources regarding the same patient and same event.

(b) A medical device report that would otherwise be required under this section is not required if:

- (1) The user facility or manufacturer determines that the information received is erroneous in that a device-related adverse event did not occur. Documentation of such reports shall be retained in MDR files for time periods specified in §803.18.
- (2) The manufacturer determines that the device was manufactured by another manufacturer. Any reportable event information that is erroneously sent to a manufacturer shall be forwarded to FDA, with a cover letter explaining that the device in question was not manufactured by that firm.

Subpart C—User Facility Reporting Requirements

§803.30 Individual adverse event reports; user facilities.

- (a) Reporting standard. A user facility shall submit the following reports to the manufacturer or to FDA, or both, as specified below:
- (1) Reports of death. Whenever a user facility receives or otherwise becomes aware of information, from any source, that reasonably suggests that a device has or may have caused or contributed to the death of a patient of the facility, the facility shall as soon as practicable, but not later than 10 work days after becoming aware of the information, report the information required by §803.32 to FDA, on FDA Form 3500A, or an electronic equivalent as approved under §803.14, and if the identity of the manufacturer is known, to the device manufacturer.
- (2) Reports of serious injury. Whenever a user facility receives or otherwise becomes aware of information, from any source, that reasonably suggests that a device has or may have caused or contributed to a serious injury to a patient of the facility, the facility shall, as soon as practicable but not later than 10 work days after becoming aware of the information, report the information required by §803.32, on FDA Form 3500A or electronic equivalent, as approved under §803.14, to the